

PerfeCta® qPCR SuperMix

Cat. No. 95050-100	Size: 100 x 50- μ L reactions (2 x 1.25 mL)
95050-500	500 x 50- μ L reactions (10 x 1.25 mL)
95050-02K	2000 x 50- μ L reactions (1 x 50 mL)

Store at -20°C

Description

PerfeCta qPCR SuperMix is a 2X concentrated, ready-to-use reaction cocktail that contains all components, except primers, probe(s), and template for real-time quantitative PCR systems that do not require an internal reference dye. The proprietary buffer and stabilizers have been specifically optimized to deliver maximum PCR efficiency, sensitivity, and robust fluorescent signal with TaqMan® or TaqMan MGB probe chemistry. The enhanced specificity of this supermix suppresses cross-reactivity between homologous sequences, improving detection and discrimination in SNP applications. A key component of this supermix is AccuStart™ Taq DNA polymerase, which contains monoclonal antibodies that bind to the polymerase and keep it inactive prior to the initial PCR denaturation step. Upon heat activation (2 minutes at 95°C), the antibodies denature irreversibly, releasing fully active, unmodified Taq DNA polymerase. This enables specific and efficient primer extension with the convenience of room temperature reaction assembly.

Instrument Compatibility

Different real-time PCR systems employ different strategies for the normalization of fluorescent signals and correction of well-to-well optical variations. It is critical to match the appropriate qPCR reagent to your specific instrument. PerfeCta qPCR SuperMix does not contain an internal reference dye. Please consult the following table, or visit our web site at www.quantabio.com to find the optimal kit for your instrument platform.

Reagent	Cat Nos	Compatible Real-Time PCR Systems
PerfeCta qPCR SuperMix, ROX	95051-100, 95051-500, 95051-02K	Applied Biosystems 7000, 7300, 7700, 7900, 7900HT, 7900HT Fast, StepOne™, StepOnePlus™
PerfeCta qPCR SuperMix, Low ROX	95052-100, 95052-500, 95052-02K	Applied Biosystems 7500, 7500 Fast, ViiA™ 7 Stratagene MX4000™, MX3005P™, MX3000P™
PerfeCta qPCR SuperMix	95050-100, 95050-500, 95050-02K	Bio-Rad CFX96™, CFX384™, iCycler iQ®, iQ™5, MyiQ™ Opticon™, MiniOpticon™, Chromo4™ Cepheid Smart Cycler®; Qiagen/Corbett Rotor-Gene® Eppendorf Mastercycler® ep realplex Roche Applied Science LightCycler® 480

Components

PerfeCta qPCR SuperMix (2X): 2X reaction buffer containing optimized concentrations of MgCl₂, dNTPs (dATP, dCTP, dGTP, dTTP), AccuStart Taq DNA Polymerase, and stabilizers.

Storage and Stability

PerfeCta qPCR SuperMix is stable for 1 year when stored in a constant temperature freezer at -20°C. For convenience, it may be stored unfrozen at +2 to +8°C for up to 6 months.

Repeated freezing and thawing of the supermix is not recommended. However, the product demonstrated no loss of performance after 20 freeze-thaw cycles or 2 months at +20°C.

Guidelines for qPCR:

- The design of highly specific primers and probes is a critical parameter for successful real-time PCR. The use of computer aided primer design programs is encouraged in order to minimize the potential for internal secondary structure and complementation at 3'-ends within each primer, the primer pair, and primer/probe combinations. PerfeCta qPCR SuperMix can readily amplify fragments between 400 and 500 bp; however, for best results, amplicon size should be limited to 65 - 200 bp. Optimal results may require titration of primer concentration between 100 and 900 nM. A final concentration of 300 nM each primer and 100 to 250 nM probe is effective for most applications. However, increasing the concentration of the primer that initiates synthesis of the target strand that is complementary to the probe can improve fluorescent signal for some primer/probe systems.

Guidelines for qPCR continued:

- Preparation of a reaction cocktail is recommended to reduce pipetting errors and maximize assay precision. Assemble the reaction cocktail with all required components except sample template (genomic DNA or cDNA) and dispense equal aliquots into each reaction tube. Add the DNA template to each reaction as the final step. Addition of samples as 5 to 10- μ L volumes will improve assay precision
- Suggested input quantities of template are: cDNA corresponding to 1 pg to 1 μ g of total RNA; 100 pg to 1 μ g genomic DNA
- After sealing each reaction, vortex gently to mix contents. Centrifuge briefly to collect components at the bottom of the reaction tube.

Reaction Assembly

Component	Volume for 50- μ L rxn.	Final Concentration
PerfeCTa qPCR SuperMix (2X)	25 μ L	1x
Forward primer	variable	100 – 900 nM
Reverse primer	variable	100 – 900 nM
Probe	variable	100 – 250 nM
Nuclease-free water	variable	
Template	5 – 10 μ L	variable
Final Volume (μ L)	50 μ L	

Note: For smaller reaction volumes (i.e. 25- μ L reactions), scale all components proportionally.

Reaction Protocol

Incubate complete reaction mix in a real-time thermal detection system as follows:

Initial denaturation:	95°C, 2 to 3 min
PCR cycling (30-45 cycles:)	95°C, 10 to 15 s
	55 – 65°C, 30 to 45 s (collect and analyze data)

Full activation of AccuStart Taq DNA polymerase occurs within 30 seconds at 95°C. Initial denaturation times greater than 3 minutes are usually not required. However, amplification of genomic DNA or supercoiled plasmid DNA targets may benefit from a prolonged initial denaturation step (5-10 min) to fully denature and fragment the template. This minimizes the potential for renaturation of long fragments and/or repetitive sequence regions that can impair replication of the target sequence by the PCR process.

Some primer sets may require a 3-step cycling protocol for optimal performance. Optimal annealing temperature and time may need to be empirically determined for any given primer set. A 68°C extension step of 30 seconds is suitable for most applications. Amplicons greater than 200 bp may require longer extension times.

Quality Control

Kit components are free of contaminating DNase and RNase. PerfeCTa qPCR SuperMix is functionally tested in qPCR. Kinetic analysis must demonstrate linear resolution over six orders of dynamic range ($r^2 > 0.995$) and a PCR efficiency $> 90\%$.

Limited Label Licenses

Use of this product is covered by one or more of the following US patents and corresponding patent claims outside the US: 5,804,375, 5,538,848, 5,723,591, 5,876,930, 6,030,787 and 6,258,569. The purchase of this product includes a limited, non-transferable immunity from suit under the foregoing patent claims for using only this amount of product for the purchaser's own internal research. No right under any other patent claim and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. Diagnostic uses under Roche patents require a separate license from Roche. Further information on purchasing licenses may be obtained from the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

Licensed to Quanta BioSciences, under U.S. Patent Nos. 5,338,671, 5,587,287, and foreign equivalents for use in research only.

PerfeCTa, and AccuStart are trademarks of Quanta BioSciences Inc. TaqMan is a registered trademark of Roche Molecular Systems, Inc. LightCycler is a registered Trademark of Roche. Applied Biosystems, StepOne, StepOnePlus, ViiA, and ROX are trademarks Life Technologies Corporation. Stratagene, MX3000P, MX3005P and MX4000 are trademarks of Agilent Technologies, Inc. Mastercycler is a trademark of Eppendorf. Rotor-Gene is a registered trademark of Qiagen GmbH. SmartCycler is a trademark of Cepheid. CFX96, CFX384, iCycler iQ, iQ5, MyiQ, Opticon, MiniOpticon and Chromo4 are trademarks of Bio-Rad Laboratories.